



The European Medicines Agency
Inspections

London, 20 September 2007
EMEA/INS/GCP/197220/2005

Procedure no: INS/GCP/3/II

<p style="text-align: center;">ANNEX II</p> <p style="text-align: center;">TO PROCEDURE FOR CONDUCTING GCP INSPECTIONS REQUESTED BY THE EMEA:</p> <p style="text-align: center;">CLINICAL LABORATORIES</p>

<p style="text-align: center;">GCP Inspectors Working Group</p>
--

Applies to: EMEA, EU/EEA Inspectorates	
Summary of scope: This procedure compiles the main aspects that are to be verified at a clinical laboratory during a GCP inspection requested by the EMEA	
Keywords: GCP Inspection, Clinical Laboratory	Public
Supersedes: N/A	

Finalisation	Date
Adoption by GCP Inspectors Working Group	5 September 2007

1	INTRODUCTION	3
2	GENERAL ASPECTS	3
2.1	BACKGROUND	3
2.2	ORGANISATION AND PERSONNEL	3
2.3	CONTRACTUAL ARRANGEMENTS	3
2.4	FACILITIES/ PREMISES	3
2.5	APPARATUS/ EQUIPMENT, MATERIALS, REAGENTS	4
3	TRIAL RELATED ASPECTS.	4
3.1	HANDLING OF SAMPLES	4
3.1.1	Pre-examination	4
3.1.2	Examination	4
3.1.3	Post-examination	4
3.2	MATERIAL AND METHODS	4
4	REPORTING	4
5	QUALITY ASSURANCE	5
6	REFERENCES	5

1. INTRODUCTION

This procedure may be applied to the inspection of laboratories involved in clinical trials, e.g. analytical chemistry, clinical biochemistry, haematology, microbiology, histopathology, cytology, genetics.

As there already is a large volume of guidelines and other documentation available on the inspection applicable to laboratories this procedure is merely presenting a general outline of the elements that have to be taken into account when inspecting such laboratories.

The following aspects should be checked during an inspection:

2. GENERAL ASPECTS

2.1. Background

- 2.1.1. Scope of work and responsibilities.
- 2.1.2. Accreditation status of the laboratory (the methods) e.g. GLP, GMP, ISO, EN.
 - Fulfilment of national requirements of accreditation.
 - Relevance of accreditation in the context of clinical trial(s).
- 2.1.3. Proportion of work in connection to clinical trials.

2.2. Organisation and Personnel

- 2.2.1. Organisation charts (facility management and scientific organisation charts).
- 2.2.2. Systems for QA and QC, including programmes for internal audits.
- 2.2.3. SOP system (distribution, availability including holidays etc., audit-trail, clinical trials, archiving etc).
- 2.2.4. Disaster plans, e.g. handling of defective equipment and consequences.
- 2.2.5. Staff – qualification, responsibilities, experience, availability, training programmes, training records, CV.

2.3. Contractual arrangements

- 2.3.1. Procedures for e.g. contracts and sub-contracts, protocol, protocol amendments, definition of source data, agreements for reporting.
- 2.3.2. Methods and procedures (including sample handling).
- 2.3.3. Agreed access and availability for monitoring, audit and inspection.
- 2.3.4. Data recording, handling and archiving.
- 2.3.5. Security and protection of subject confidentiality.

2.4. Facilities/ Premises

- 2.4.1. Suitability and adequacy of premises – e.g. adequate degree of separation of work areas to avoid mix-ups, contamination and interference.
- 2.4.2. Environmental conditions, e.g. temperature, airflow and air pressure, microbiological contamination.
- 2.4.3. Security and safety, e.g. fire, water and pest control.

2.5. Apparatus/ Equipment, Materials, Reagents

- 2.5.1. Apparatus available, in good working order and complies with relevant specifications.
- 2.5.2. Quality of general supplies including tap water, analytical water, gases etc.
- 2.5.3. Records of operation, maintenance, justification and calibration. Records of the validation for the methods used for the measuring equipment and apparatus (including computerised systems). Log books.
- 2.5.4. Materials and reagents are prepared, labelled and stored under appropriate conditions and adherence to expiry dates. Labels for reagents indicate their identity, source, concentration and expiry dates.
- 2.5.5. Apparatus and materials used do not alter to any appreciable extent the samples.
- 2.5.6. Definition of source data and source documents, retrieval and archiving. Data generated in automatic systems e.g. listings, graphs, record traces or computer printouts are archived.

3. TRIAL RELATED ASPECTS.

As part of the inspection all aspects applicable to the clinical trial, as listed under section 2 should also be inspected.

3.1. Handling of samples

- 3.1.1. Pre-examination
 - Samples obtained from subjects in the clinical laboratory, (date and time), identification, labelling, conditions, preparation, storage.
 - Documentation of receipt (date and time), identification, condition, re-labelling and storage of samples by identifiable person.
 - Procedures for acceptance or rejection of samples for analysis.
 - Aliquoting and distribution for examination.
- 3.1.2. Examination
 - Compliance with protocol and specified test methods.
 - Traceability and identification of samples and controls.
 - Recording of data and acceptance and release of results.
 - Handling of non-conformance, repeat analysis / re-analysis, and results within critical / alert ranges.
 - Competence, training and experience of personnel.
- 3.1.3. Procedures for disaster recovery.
- 3.1.4. Post-examination
 - Storage (anonymisation, decoding), retrieval and destruction of samples.

3.2. Material and methods

- 3.2.1. Material and methods according to the specification stated in the protocol / contract and/or required according to Ph Eur
- 3.2.2. Validation status of the methods, appropriately setting of limits of detection / quantification, precision/accuracy, known inferences and specific control measures.
- 3.2.3. Participation in external control programmes, if applicable.

4. REPORTING

Various systems for reporting of results may be required according to the protocol/contract e.g. report per sample (i.e. for immediate consideration in medical care of the subject) or on an integrated basis (i.e. to be used in the trial report). This will affect the procedures used by the laboratory and the inspection.

- 4.1. Procedures for reporting and evaluation of results and for data transfer.**
- 4.2. Systems for alerting results that are unexpected and/or significant deviations from pre-specified limits.**
- 4.3. Transcription of raw data into the report**
 - 4.3.1. Identification of laboratory,
 - 4.3.2. Unique identification and localisation of the subject,
 - 4.3.3. Identification of investigator,
 - 4.3.4. Date and time of sample collection, and time of receipt,
 - 4.3.5. Date and time of examination and release of report,
 - 4.3.6. Source of primary sample type and any comments of its quality,
 - 4.3.7. Description of the examination and of its results,
 - 4.3.8. If applicable, detection limit, uncertainty of each measurements, and reference intervals,
 - 4.3.9. Where appropriate, interpretation of results and other comments,
 - 4.3.10. Identification of the person releasing the report.
- 4.4. Attribution of review and release of the report(s) to responsible personnel.**
- 4.5. Procedures for alterations and amendments of reports.**
- 4.6. Procedures for complaints and corrective actions.**

5. QUALITY ASSURANCE

- 5.1. Integrity of data reported by internal QA/QC and /or sponsor's QA/QC personnel, (audit certificate)**

6. REFERENCES

- See "Principal documents taken into account for the preparation of procedures for GCP inspections requested by the EMEA".